

**Description**

Blue colored gel-type substance containing 37% phosphoric acid that etches enamel and dentine to promote adhesion of primer/bonding agent adhesives to tooth structure and restorative materials.

**Applications**

**A. Preparations before use**

- 1) Read the instructions for use and familiarize with the application method and precautions before use.
- 2) Check the condition of product including expiry date, presence of foreign matters and impurities, and deformation.
- 3) Open the product after having thoroughly assessed the state of the surrounding environment.
- 4) A rubber dam should be used to protect the patient's mouth

**B. Recommended Procedure**

- 1) Prepare tooth and cleanse the enamel surface to be etched. Rinse and dry thoroughly.
- 2) Apply DiaEtch onto the tooth, recommended etching time is 15 seconds.
- 3) Rinse off DiaEtch with large amount of water to ensure there is no residue left.
- 4) Take precaution against any saliva contamination of the treated surface as you proceed to the next procedure.
- 5) Remove the disposable tip and tightly cap the syringe immediately after use.

**C. Storage**

- Avoid high temperature and direct sunlight.
- Store in a dry and cool location (2~27°C)
- Keep it out of reach of children.
- Do not store near open flame.

**Cautions**

**A. Contraindication**

- Do not use on patients with past history of hypersensitivity and allergy to any of this product's components.

**B. General Precautions**

- This product should only be used by a dentist or a dental professional

- Do not use this product for purposes other than the prescribed purpose of use.
- Wear acid resistant protective wears (i.e. protective glasses, mask and gloves) when using the product.
- Tightly close and cover the product with the provided cap immediately after use.
- Do not use products that have expired.

**C. Application Precautions**

- Take precaution against product coming in contact with oral mucous membrane, skin and eyes. In case of contact, wipe off with alcohol soaked cotton ball immediately and wash off with large amount of running water. If needed, seek medical assistance.
- Only use the disposable tip designed for DiaEtch
- Securely attach the disposable tip onto the syringe before procedure.
- Start the flow slowly. DO NOT push the plunger forcefully. Use controlled light force to slowly and gently push the plunger to dispense product.
- Never reuse the disposable tip on other patients. Dispose the used tip after use.

**Shelf Life**

2 years from date of manufacture

**Symbols**



**Warranty**

DiaDent Group International will replace product proven to have defects. DiaDent Group International shall not be responsible for any direct or indirect losses or damages arising from the use of or failure to use such product. User must decide the appropriateness of the intended use of the product prior to the actual use and must bear all responsibilities, risks and consequences associated with such decision.

**Description**

Blue colored gel-type substance containing 37% phosphoric acid that etches enamel and dentine to promote adhesion of primer/bonding agent adhesives to tooth structure and restorative materials.

**Applications**

**A. Preparations before use**

- 1) Read the instructions for use and familiarize with the application method and precautions before use.
- 2) Check the condition of product including expiry date, presence of foreign matters and impurities, and deformation.
- 3) Open the product after having thoroughly assessed the state of the surrounding environment.
- 4) A rubber dam should be used to protect the patient's mouth

**B. Recommended Procedure**

- 1) Prepare tooth and cleanse the enamel surface to be etched. Rinse and dry thoroughly.
- 2) Apply DiaEtch onto the tooth, recommended etching time is 15 seconds.
- 3) Rinse off DiaEtch with large amount of water to ensure there is no residue left.
- 4) Take precaution against any saliva contamination of the treated surface as you proceed to the next procedure.
- 5) Remove the disposable tip and tightly cap the syringe immediately after use.

**C. Storage**

- Avoid high temperature and direct sunlight.
- Store in a dry and cool location (2~27°C)
- Keep it out of reach of children.
- Do not store near open flame.

**Cautions**

**A. Contraindication**

- Do not use on patients with past history of hypersensitivity and allergy to any of this product's components.

**B. General Precautions**

- This product should only be used by a dentist or a dental professional

- Do not use this product for purposes other than the prescribed purpose of use.
- Wear acid resistant protective wears (i.e. protective glasses, mask and gloves) when using the product.
- Tightly close and cover the product with the provided cap immediately after use.
- Do not use products that have expired.

**C. Application Precautions**

- Take precaution against product coming in contact with oral mucous membrane, skin and eyes. In case of contact, wipe off with alcohol soaked cotton ball immediately and wash off with large amount of running water. If needed, seek medical assistance.
- Only use the disposable tip designed for DiaEtch
- Securely attach the disposable tip onto the syringe before procedure.
- Start the flow slowly. DO NOT push the plunger forcefully. Use controlled light force to slowly and gently push the plunger to dispense product.
- Never reuse the disposable tip on other patients. Dispose the used tip after use.

**Shelf Life**

2 years from date of manufacture

**Symbols**



**Warranty**

DiaDent Group International will replace product proven to have defects. DiaDent Group International shall not be responsible for any direct or indirect losses or damages arising from the use of or failure to use such product. User must decide the appropriateness of the intended use of the product prior to the actual use and must bear all responsibilities, risks and consequences associated with such decision.

**DiaDent Group International**

**DiaDent Group International**  
16, Osongsaengmyeong 4-ro, Osong-eup,  
Heungdeok-gu, Cheongju-si,  
Chungcheongbuk-do, 28161, Korea  
Tel : 82-43-266-2315  
Fax: 82-43-262-8658  
www.diaindent.co.kr  
diadent@diaindent.co.kr

**DiaDent Group International Inc.**  
11-3871 North Fraser Way, Burnaby  
B.C., Canada, V5J 5G6  
Tel : 1-604-451-8851  
Fax: 1-604-451-8865  
www.diaindent.com  
diadent@diaindent.com

**DiaDent Europe B.V.**  
Antennestraat 70, 1322AS  
Almere, the Netherlands  
Tel : 31-36-549-8607  
Fax: 31-36-536-7317  
www.diaident.europa.com  
diadent@diaidenteuropa.com



**DiaDent Group International**

**DiaDent Group International**  
16, Osongsaengmyeong 4-ro, Osong-eup,  
Heungdeok-gu, Cheongju-si,  
Chungcheongbuk-do, 28161, Korea  
Tel : 82-43-266-2315  
Fax: 82-43-262-8658  
www.diaindent.co.kr  
diadent@diaindent.co.kr

**DiaDent Group International Inc.**  
11-3871 North Fraser Way, Burnaby  
B.C., Canada, V5J 5G6  
Tel : 1-604-451-8851  
Fax: 1-604-451-8865  
www.diaindent.com  
diadent@diaindent.com

**DiaDent Europe B.V.**  
Antennestraat 70, 1322AS  
Almere, the Netherlands  
Tel : 31-36-549-8607  
Fax: 31-36-536-7317  
www.diaident.europa.com  
diadent@diaidenteuropa.com

